



calibration
(1 : 2)



measurement
(1 : 2)



gauge 0641-08



type: Aerosol
range: 0 - 25 mm
range of indication: 0,01 mm
digital step: 0,01 mm
scale interval: 0,7 - 1,3 N
measuring force: ca. 330 g
weight:

acceptable deviations
VDI/VDE/DGQ 2618 Bl. 12.1/13.1)
max. permissible errors "G": 0,03 mm
repeatability "r": 0,02 mm
reference temperature: 20° C
operation temperature: 10° C - 30° C
storage temperature: -10° C - 50° C

scale: 1:1 (1:2)
drawing-nr.: IDW-AE1210-DB
revision status: 2123 / JL
revision date: 23.09.2015